

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111518\
 Data File : FG002969.D
 Signal(s) : FID1A.ch
 Acq On : 15 Nov 2018 10:18
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 15 16:00:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
 Quant Title :
 QLast Update : Sat Oct 27 02:13:37 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.749	5361618	48.786 ug/mlm
Target Compounds			
2) N-DECANE	4.459	7062666	55.375 ug/ml
3) N-DODECANE	6.488	7009365	53.777 ug/ml
4) N-TETRADECANE	8.287	6932161	52.427 ug/ml
5) N-HEXADECANE	9.891	6817478	51.197 ug/ml
6) N-OCTADECANE	11.335	6695305	50.612 ug/ml
7) N-EICOSANE	12.648	6497747	50.124 ug/ml
8) N-DOCOSANE	13.848	6360156	49.962 ug/ml
10) N-TETRACOSANE	14.954	6210104	49.710 ug/ml
11) N-HEXACOSANE	15.976	6241892	50.646 ug/ml
12) N-OCTACOSANE	16.926	5995835	49.297 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111518\
Data File : FG002969.D
Signal(s) : FID1A.ch
Acq On : 15 Nov 2018 10:18
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 15 16:00:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
Quant Title :
QLast Update : Sat Oct 27 02:13:37 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

