

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111521\
 Data File : FG009080.D
 Signal(s) : FID1A.ch
 Acq On : 15 Nov 2021 20:10
 Operator : AJ\MA
 Sample : PB140763BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB140763BS

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/16/2021
 Supervised By :mohammad ahmed 11/16/2021

Integration File: autoint1.e
 Quant Time: Nov 16 00:59:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
 Quant Title :
 QLast Update : Tue Nov 02 01:56:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.752	2126949	18.428 ug/ml
Target Compounds			
1) N-OCTANE	2.334	2737874	20.303 ug/ml
2) N-DECANE	4.438	2742052	20.279 ug/ml
3) N-DODECANE	6.466	2721385	20.042 ug/ml
4) N-TETRADECANE	8.271	2637243	19.856 ug/ml
5) N-HEXADECANE	9.878	2568587	19.376 ug/ml
6) N-OCTADECANE	11.327	2617262	19.088 ug/ml
7) N-EICOSANE	12.643	2563755	18.744 ug/ml
8) N-DOCOSANE	13.846	2458157	18.368 ug/ml
10) N-TETRACOSANE	14.957	2404906	18.207 ug/ml
11) N-HEXACOSANE	15.982	2324819	18.260 ug/ml
12) N-OCTACOSANE	16.936	2292026	18.644 ug/ml
13) N-TRIACONTANE	17.828	2294273	19.595 ug/ml
14) N-DOTRIACONTANE	18.664	2259224	20.751 ug/ml
15) N-TETRATRIACONTANE	19.451	2218369	21.838 ug/ml
16) N-HEXATRIACONTANE	20.196	2144862	22.740 ug/mlm
17) N-OCTATRIACONTANE	20.952	2043446	23.095 ug/mlm
18) N-TETRACONTANE	21.898	2000839	22.647 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111521\
Data File : FG009080.D
Signal(s) : FID1A.ch
Acq On : 15 Nov 2021 20:10
Operator : AJ\MA
Sample : PB140763BS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB140763BS

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/16/2021
Supervised By :mohammad ahmed 11/16/2021

Integration File: autoint1.e
Quant Time: Nov 16 00:59:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG110121.M
Quant Title :
QLast Update : Tue Nov 02 01:56:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

