

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111618\
 Data File : FG002982.D
 Signal(s) : FID1A.ch
 Acq On : 16 Nov 2018 12:09
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 16 23:51:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
 Quant Title :
 QLast Update : Sat Oct 27 02:13:37 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.749	5265491	47.911 ug/ml
Target Compounds			
1) N-OCTANE	2.337	6636851	53.087 ug/ml
2) N-DECANE	4.457	7118741	55.815 ug/ml
3) N-DODECANE	6.487	7119384	54.621 ug/ml
4) N-TETRADECANE	8.286	7045494	53.284 ug/ml
5) N-HEXADECANE	9.891	6924972	52.004 ug/ml
6) N-OCTADECANE	11.335	6757695	51.083 ug/ml
7) N-EICOSANE	12.648	6496171	50.112 ug/ml
8) N-DOCOSANE	13.849	6259593	49.172 ug/ml
10) N-TETRACOSANE	14.955	6046472	48.400 ug/ml
11) N-HEXACOSANE	15.977	5928153	48.100 ug/ml
12) N-OCTACOSANE	16.928	5931874	48.771 ug/ml
13) N-TRIACONTANE	17.818	6047171	50.397 ug/ml
14) N-DOTRIACONTANE	18.651	5986926	51.934 ug/ml
15) N-TETRATRIACONTANE	19.436	5877714	52.855 ug/ml
16) N-HEXATRIACONTANE	20.178	5629996	53.467 ug/ml
17) N-OCTATRIACONTANE	20.940	5209294	54.729 ug/ml
18) N-TETRACONTANE	21.893	5193666	52.582 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG111618\
Data File : FG002982.D
Signal(s) : FID1A.ch
Acq On : 16 Nov 2018 12:09
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 16 23:51:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102618.M
Quant Title :
QLast Update : Sat Oct 27 02:13:37 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

