

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112119\
 Data File : FG006260.D
 Signal(s) : FID1A.ch
 Acq On : 21 Nov 2019 13:20
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 22 00:59:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG103019.M
 Quant Title :
 QLast Update : Thu Oct 31 03:29:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.480	3774865	47.534 ug/ml
Target Compounds			
1) N-OCTANE	2.337	4350933	44.286 ug/ml
2) N-DECANE	4.395	4723066	46.947 ug/ml
3) N-DODECANE	6.379	4838468	48.011 ug/ml
4) N-TETRADECANE	8.140	4819122	48.267 ug/ml
5) N-HEXADECANE	9.711	4789576	48.231 ug/ml
6) N-OCTADECANE	11.128	4644244	47.797 ug/ml
7) N-EICOSANE	12.416	4549621	47.365 ug/ml
8) N-DOCOSANE	13.594	4400105	47.266 ug/ml
10) N-TETRACOSANE	14.680	4293845	47.276 ug/ml
11) N-HEXACOSANE	15.685	4210233	47.451 ug/ml
12) N-OCTACOSANE	16.620	4151164	47.178 ug/ml
13) N-TRIACONTANE	17.493	4112215	46.729 ug/ml
14) N-DOTRIACONTANE	18.311	4030976	47.188 ug/ml
15) N-TETRATRIACONTANE	19.083	4019608	52.678 ug/ml
16) N-HEXATRIACONTANE	19.812	3826395	48.008 ug/ml
17) N-OCTATRIACONTANE	20.506	3550170	57.023 ug/ml
18) N-TETRACONTANE	21.298	3546562	53.358 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112119\
Data File : FG006260.D
Signal(s) : FID1A.ch
Acq On : 21 Nov 2019 13:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 22 00:59:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG103019.M
Quant Title :
QLast Update : Thu Oct 31 03:29:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

