

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112122\
 Data File : FG011127.D
 Signal(s) : FID1A.ch
 Acq On : 21 Nov 2022 11:06
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 21 14:40:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG111422.M
 Quant Title :
 QLast Update : Mon Nov 14 17:25:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.078	5688596	51.963 ug/ml
Target Compounds			
1) N-OCTANE	2.073	5851623	48.827 ug/ml
2) N-DECANE	4.590	6385269	51.527 ug/ml
3) N-DODECANE	6.763	6466402	51.462 ug/ml
4) N-TETRADECANE	8.597	6491577	51.439 ug/ml
5) N-HEXADECANE	10.210	6595020	51.629 ug/ml
6) N-OCTADECANE	11.660	6695270	51.880 ug/ml
7) N-EICOSANE	12.974	6587501	52.009 ug/ml
8) N-DOCOSANE	14.176	6531862	52.001 ug/ml
10) N-TETRACOSANE	15.283	6496453	51.858 ug/ml
11) N-HEXACOSANE	16.307	6367379	51.576 ug/ml
12) N-OCTACOSANE	17.258	6186578	51.628 ug/ml
13) N-TRIACONTANE	18.149	6031582	51.325 ug/ml
14) N-DOTRIACONTANE	18.985	5793586	51.386 ug/ml
15) N-TETRATRIACONTANE	19.771	5504258	51.769 ug/ml
16) N-HEXATRIACONTANE	20.513	5338833	53.583 ug/ml
17) N-OCTATRIACONTANE	21.270	5483244	54.950 ug/ml
18) N-TETRACONTANE	22.216	5493276	53.861 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112122\
Data File : FG011127.D
Signal(s) : FID1A.ch
Acq On : 21 Nov 2022 11:06
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 21 14:40:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG111422.M
Quant Title :
QLast Update : Mon Nov 14 17:25:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

