

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112122\
 Data File : FG011131.D
 Signal(s) : FID1A.ch
 Acq On : 21 Nov 2022 13:06
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 21 14:41:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG111422.M
 Quant Title :
 QLast Update : Mon Nov 14 17:25:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.076	5676067	51.848 ug/ml
Target Compounds			
1) N-OCTANE	2.078	5743805	47.927 ug/ml
2) N-DECANE	4.596	6277462	50.657 ug/ml
3) N-DODECANE	6.766	6349900	50.535 ug/ml
4) N-TETRADECANE	8.598	6370514	50.480 ug/ml
5) N-HEXADECANE	10.210	6471904	50.666 ug/ml
6) N-OCTADECANE	11.658	6586691	51.039 ug/ml
7) N-EICOSANE	12.972	6505665	51.363 ug/ml
8) N-DOCOSANE	14.174	6487734	51.649 ug/ml
10) N-TETRACOSANE	15.281	6489856	51.805 ug/ml
11) N-HEXACOSANE	16.304	6403065	51.865 ug/ml
12) N-OCTACOSANE	17.256	6243303	52.101 ug/ml
13) N-TRIACONTANE	18.146	6105673	51.956 ug/ml
14) N-DOTRIACONTANE	18.982	5856896	51.947 ug/ml
15) N-TETRATRIACONTANE	19.767	5506172	51.787 ug/ml
16) N-HEXATRIACONTANE	20.509	5239606	52.587 ug/ml
17) N-OCTATRIACONTANE	21.265	5399317	54.109 ug/ml
18) N-TETRACONTANE	22.211	5561213	54.527 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112122\
Data File : FG011131.D
Signal(s) : FID1A.ch
Acq On : 21 Nov 2022 13:06
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 21 14:41:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG111422.M
Quant Title :
QLast Update : Mon Nov 14 17:25:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

