

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112219\
 Data File : FG006275.D
 Signal(s) : FID1A.ch
 Acq On : 22 Nov 2019 10:59
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 22 21:39:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG103019.M
 Quant Title :
 QLast Update : Thu Oct 31 03:29:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.488	3823830	48.151 ug/ml
Target Compounds			
1) N-OCTANE	2.338	4450330	45.298 ug/ml
2) N-DECANE	4.398	4926011	48.965 ug/ml
3) N-DODECANE	6.384	5049141	50.102 ug/ml
4) N-TETRADECANE	8.146	5008195	50.161 ug/ml
5) N-HEXADECANE	9.718	4937960	49.725 ug/ml
6) N-OCTADECANE	11.136	4752732	48.914 ug/ml
7) N-EICOSANE	12.425	4629753	48.199 ug/ml
8) N-DOCOSANE	13.604	4472166	48.040 ug/ml
10) N-TETRACOSANE	14.689	4383106	48.259 ug/ml
11) N-HEXACOSANE	15.694	4324921	48.743 ug/ml
12) N-OCTACOSANE	16.629	4300095	48.871 ug/ml
13) N-TRIACONTANE	17.504	4290241	48.752 ug/ml
14) N-DOTRIACONTANE	18.324	4214288	49.334 ug/ml
15) N-TETRATRIACONTANE	19.095	4191649	54.932 ug/ml
16) N-HEXATRIACONTANE	19.824	4004238	50.239 ug/ml
17) N-OCTATRIACONTANE	20.521	3725949	59.846 ug/ml
18) N-TETRACONTANE	21.319	3695710	55.602 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG112219\
Data File : FG006275.D
Signal(s) : FID1A.ch
Acq On : 22 Nov 2019 10:59
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 22 21:39:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG103019.M
Quant Title :
QLast Update : Thu Oct 31 03:29:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

