

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120123\
 Data File : FG013223.D
 Signal(s) : FID1A.ch
 Acq On : 01 Dec 2023 13:36
 Operator : YP\AJ
 Sample : PB157502BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 01 22:17:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG111023.M
 Quant Title :
 QLast Update : Fri Nov 10 12:13:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.087	1830879	17.862 ug/ml
Target Compounds			
1) N-OCTANE	2.069	1941089	16.651 ug/ml
2) N-DECANE	4.593	2017087	17.084 ug/ml
3) N-DODECANE	6.769	2000626	17.123 ug/ml
4) N-TETRADECANE	8.604	1959044	17.352 ug/ml
5) N-HEXADECANE	10.218	2007054	17.521 ug/ml
6) N-OCTADECANE	11.667	2006329	17.576 ug/ml
7) N-EICOSANE	12.982	1991326	17.644 ug/ml
8) N-DOCOSANE	14.185	1930533	17.777 ug/ml
10) N-TETRACOSANE	15.292	1998079	17.628 ug/ml
11) N-HEXACOSANE	16.317	1922961	18.124 ug/ml
12) N-OCTACOSANE	17.268	1947663	18.639 ug/ml
13) N-TRIACONTANE	18.159	2027673	19.612 ug/ml
14) N-DOTRIACONTANE	18.992	2005349	20.176 ug/ml
15) N-TETRATRIACONTANE	19.779	1735866	19.876 ug/ml
16) N-HEXATRIACONTANE	20.518	1373598	19.288 ug/ml
17) N-OCTATRIACONTANE	21.277	1235881	18.731 ug/ml
18) N-TETRACONTANE	22.234	1101974	16.197 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120123\
Data File : FG013223.D
Signal(s) : FID1A.ch
Acq On : 01 Dec 2023 13:36
Operator : YP\AJ
Sample : PB157502BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 01 22:17:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG111023.M
Quant Title :
QLast Update : Fri Nov 10 12:13:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

