

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120220\
 Data File : FG007400.D
 Signal(s) : FID1A.ch
 Acq On : 02 Dec 2020 22:45
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Ankita
 12/3/2020 9:12:03 AM

Integration File: autoint1.e
 Quant Time: Dec 03 00:44:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120220.M
 Quant Title :
 QLast Update : Thu Dec 03 00:42:51 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.734	604648	4.745 ug/ml
Target Compounds			
1) N-OCTANE	2.334	703312	4.822 ug/ml
2) N-DECANE	4.437	710013	4.794 ug/ml
3) N-DODECANE	6.464	712035	4.763 ug/ml
4) N-TETRADECANE	8.266	712202	4.729 ug/ml
5) N-HEXADECANE	9.871	713978	4.759 ug/ml
6) N-OCTADECANE	11.318	725112	4.774 ug/ml
7) N-EICOSANE	12.632	723142	4.796 ug/ml
8) N-DOCOSANE	13.832	705029	4.781 ug/ml
10) N-TETRACOSANE	14.939	702390	4.789 ug/ml
11) N-HEXACOSANE	15.964	685420	4.802 ug/ml
12) N-OCTACOSANE	16.913	666357	4.843 ug/ml
13) N-TRIACONTANE	17.803	644095	4.998 ug/ml
14) N-DOTRIACONTANE	18.638	575878	5.142 ug/ml
15) N-TETRATRIACONTANE	19.421	445130	5.029 ug/ml
16) N-HEXATRIACONTANE	20.165	346116	5.216 ug/mlm
17) N-OCTATRIACONTANE	20.918	282747	5.043 ug/ml
18) N-TETRACONTANE	21.862	262433	4.910 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120220\
Data File : FG007400.D
Signal(s) : FID1A.ch
Acq On : 02 Dec 2020 22:45
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Ankita
12/3/2020 9:12:03 AM

Integration File: autoint1.e
Quant Time: Dec 03 00:44:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120220.M
Quant Title :
QLast Update : Thu Dec 03 00:42:51 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

