

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 12/03/2021

Supervised By :mohammad ahmed 12/06/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG12022
Data File : FG009129.D
Signal(s) : FID1A.ch
Acq On : 02 Dec 2021 11:01
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112321.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.334	2.300	2.395	BB	1002154	7261266	97.88%	6.477%
2	4.439	4.410	4.526	BB	810018	7418628	100.00%	6.618%
3	6.470	6.436	6.554	BB	726356	7408395	99.86%	6.609%
4	8.275	8.241	8.366	BB	694847	7302399	98.43%	6.514%
5	9.885	9.848	9.970	BB	648536	7088077	95.54%	6.323%
6	11.335	11.292	11.401	BB	624569	7047854	95.00%	6.287%
7	12.652	12.611	12.720	BB	563440	6838063	92.17%	6.100%
8	13.857	13.815	13.922	BB	515149	6496333	87.57%	5.795%
9	14.761	14.709	14.834	BB	401706	5498224	74.11%	4.905%
10	14.967	14.915	15.024	BB	483398	6224707	83.91%	5.553%
11	15.993	15.945	16.052	BB	449139	5880495	79.27%	5.246%
12	16.949	16.884	17.011	BB	436105	5647183	76.12%	5.037%
13	17.841	17.785	17.900	BB	404794	5526931	74.50%	4.930%
14	18.679	18.620	18.736	BB	391208	5434020	73.25%	4.847%
15	19.467	19.192	19.520	BB	367221	5408954	72.91%	4.825%
16	20.210	20.147	20.266	M	360856	5393398	72.70%	4.811%
17	20.971	20.852	21.041	BV	281284	5046811	68.03%	4.502%
18	21.925	21.839	22.001	PBA	212145	5181361	69.84%	4.622%
Sum of corrected areas:						112103101		

FG112321.M Fri Dec 03 09:50:05 2021