

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Rahul Chavli 12/03/2025

Supervised By :mohammad ahmed 12/05/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120125.D
Data File : FG017054.D
Signal(s) : FID1A.ch
Acq On : 03 Dec 2025 04:58
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120125.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.081	2.049	2.215	PB	586393	6099875	97.29%	6.129%
2	4.620	4.577	4.745	BB	597436	6191053	98.74%	6.220%
3	6.803	6.762	6.932	BB	606901	6213905	99.11%	6.243%
4	8.643	8.605	8.830	BB	567567	6155858	98.18%	6.185%
5	10.261	10.222	10.442	BB	529726	6067527	96.77%	6.096%
6	11.714	11.662	11.872	BB	522387	6190033	98.73%	6.219%
7	13.034	12.980	13.184	BB	507276	6269971	100.00%	6.300%
8	14.241	14.137	14.330	BB	477609	5981678	95.40%	6.010%
9	15.147	15.104	15.262	BB	371951	5199138	82.92%	5.224%
10	15.353	15.284	15.474	BB	467423	5885954	93.88%	5.914%
11	16.381	16.252	16.489	BB	421005	5593789	89.22%	5.620%
12	17.337	16.937	17.450	BB	410698	5505156	87.80%	5.531%
13	18.231	18.160	18.304	PB	400234	5462194	87.12%	5.488%
14	19.070	19.032	19.144	M	375873	5325395	84.93%	5.350%
15	19.859	19.808	19.901	M	348885	4998214	79.72%	5.022%
16	20.602	20.566	20.644	M	304886	4466329	71.23%	4.487%
17	21.381	21.335	21.436	M	210905	4000168	63.80%	4.019%
18	22.359	22.294	22.477	M	151134	3924977	62.60%	3.943%
Sum of corrected areas:							99531211	

FG120125.M Wed Dec 03 05:28:09 2025