

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120318\
 Data File : FG003134.D
 Signal(s) : FID1A.ch
 Acq On : 03 Dec 2018 10:36
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 03 14:49:45 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112818.M
 Quant Title :
 QLast Update : Thu Nov 29 04:54:55 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.716	4859283	44.589 ug/ml
Target Compounds			
1) N-OCTANE	2.303	5618039	49.150 ug/ml
2) N-DECANE	4.417	5721966	48.966 ug/ml
3) N-DODECANE	6.444	5764518	48.461 ug/ml
4) N-TETRADECANE	8.244	5770351	47.495 ug/ml
5) N-HEXADECANE	9.850	5758909	46.440 ug/ml
6) N-OCTADECANE	11.296	5738114	45.648 ug/ml
7) N-EICOSANE	12.610	5658291	45.090 ug/ml
8) N-DOCOSANE	13.812	5594964	44.739 ug/ml
10) N-TETRACOSANE	14.922	5523417	44.439 ug/ml
11) N-HEXACOSANE	15.947	5485132	44.147 ug/ml
12) N-OCTACOSANE	16.903	5461462	43.824 ug/ml
13) N-TRIACONTANE	17.795	5456441	42.880 ug/ml
14) N-DOTRIACONTANE	18.637	5307820	42.356 ug/ml
15) N-TETRATRIACONTANE	19.426	5174640	42.390 ug/mlm
16) N-HEXATRIACONTANE	20.187	5045982	42.054 ug/ml
17) N-OCTATRIACONTANE	20.952	4603066	43.417 ug/ml
18) N-TETRACONTANE	21.917	1963695	17.331 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120318\
Data File : FG003134.D
Signal(s) : FID1A.ch
Acq On : 03 Dec 2018 10:36
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 03 14:49:45 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112818.M
Quant Title :
QLast Update : Thu Nov 29 04:54:55 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

