

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120318\
 Data File : FG003141.D
 Signal(s) : FID1A.ch
 Acq On : 03 Dec 2018 15:03
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 03 15:24:45 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112818.M
 Quant Title :
 QLast Update : Thu Nov 29 04:54:55 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.716	5024081	46.101 ug/ml
Target Compounds			
1) N-OCTANE	2.304	5724045	50.078 ug/ml
2) N-DECANE	4.417	5886321	50.372 ug/ml
3) N-DODECANE	6.445	5941237	49.947 ug/ml
4) N-TETRADECANE	8.245	5960846	49.063 ug/ml
5) N-HEXADECANE	9.849	5951548	47.993 ug/ml
6) N-OCTADECANE	11.295	5931370	47.185 ug/ml
7) N-EICOSANE	12.610	5843790	46.568 ug/ml
8) N-DOCOSANE	13.813	5763943	46.091 ug/ml
10) N-TETRACOSANE	14.920	5680589	45.703 ug/ml
11) N-HEXACOSANE	15.947	5636192	45.363 ug/ml
12) N-OCTACOSANE	16.902	5608628	45.005 ug/ml
13) N-TRIACONTANE	17.796	5610536	44.091 ug/ml
14) N-DOTRIACONTANE	18.636	5460885	43.577 ug/ml
15) N-TETRATRIACONTANE	19.435	5281007	43.261 ug/mlm
16) N-HEXATRIACONTANE	20.178	5151192	42.931 ug/ml
17) N-OCTATRIACONTANE	20.955	4677635	44.120 ug/ml
18) N-TETRACONTANE	21.913	3807965	33.609 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120318\
Data File : FG003141.D
Signal(s) : FID1A.ch
Acq On : 03 Dec 2018 15:03
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 03 15:24:45 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112818.M
Quant Title :
QLast Update : Thu Nov 29 04:54:55 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

