

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120320\
 Data File : FG007413.D
 Signal(s) : FID1A.ch
 Acq On : 03 Dec 2020 16:24
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Ankita
 12/7/2020 12:15:31 PM

Integration File: autoint1.e
 Quant Time: Dec 04 00:13:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120220.M
 Quant Title :
 QLast Update : Thu Dec 03 08:50:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.737	5999806	46.791 ug/ml
Target Compounds			
1) N-OCTANE	2.332	7405902	50.153 ug/ml
2) N-DECANE	4.438	7549605	50.445 ug/ml
3) N-DODECANE	6.466	7553489	50.128 ug/ml
4) N-TETRADECANE	8.268	7518443	49.618 ug/ml
5) N-HEXADECANE	9.874	7396381	48.903 ug/ml
6) N-OCTADECANE	11.320	7405751	48.336 ug/ml
7) N-EICOSANE	12.635	7270361	47.763 ug/ml
8) N-DOCOSANE	13.835	7019521	47.207 ug/ml
10) N-TETRACOSANE	14.942	6878128	46.492 ug/ml
11) N-HEXACOSANE	15.965	6569818	45.647 ug/ml
12) N-OCTACOSANE	16.916	6223105	44.821 ug/ml
13) N-TRIACONTANE	17.805	5804967	44.393 ug/ml
14) N-DOTRIACONTANE	18.639	5129208	44.741 ug/ml
15) N-TETRATRIACONTANE	19.423	4214946	46.923 ug/ml
16) N-HEXATRIACONTANE	20.165	3894607	58.782 ug/mlm
17) N-OCTATRIACONTANE	20.919	2829598	50.054 ug/ml
18) N-TETRACONTANE	21.861	2759862	50.734 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120320\
Data File : FG007413.D
Signal(s) : FID1A.ch
Acq On : 03 Dec 2020 16:24
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
12/7/2020 12:15:31 PM

Integration File: autoint1.e
Quant Time: Dec 04 00:13:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120220.M
Quant Title :
QLast Update : Thu Dec 03 08:50:10 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

