

rtres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 12/06/2021
Supervised By :mohammad ahmed 12/06/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG12032
Data File : FG009138.D
Signal(s) : FID1A.ch
Acq On : 03 Dec 2021 10:57
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112321.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.332	2.300	2.395	BB	938810	6903598	98.12%	6.626%
2	4.437	4.407	4.520	BB	774798	7036123	100.00%	6.753%
3	6.468	6.430	6.551	BB	693685	7028908	99.90%	6.747%
4	8.273	8.229	8.360	BB	643613	6918869	98.33%	6.641%
5	9.883	9.835	9.967	BB	598684	6707753	95.33%	6.438%
6	11.333	11.289	11.406	BB	582637	6679752	94.94%	6.411%
7	12.650	12.601	12.723	BB	550609	6478690	92.08%	6.218%
8	13.856	13.818	13.927	BB	505763	6157866	87.52%	5.910%
9	14.759	14.721	14.835	BB	380246	5247443	74.58%	5.037%
10	14.964	14.915	15.024	BB	472098	5903632	83.90%	5.666%
11	15.992	15.944	16.054	BB	427168	5572364	79.20%	5.348%
12	16.947	16.895	17.010	BB	401483	5321643	75.63%	5.108%
13	17.839	17.786	17.906	BB	377943	5164402	73.40%	4.957%
14	18.676	18.624	18.752	BB	355206	5024466	71.41%	4.823%
15	19.464	19.421	19.519	M	347915	4873704	69.27%	4.678%
16	20.207	20.075	20.275	BB	316824	4556840	64.76%	4.374%
17	20.967	20.902	21.048	BB	235891	4317131	61.36%	4.144%
18	21.919	21.837	22.000	BBA	169913	4292506	61.01%	4.120%
Sum of corrected areas:						104185690		

FG112321.M Mon Dec 06 10:18:59 2021