

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120618\
 Data File : FG003171.D
 Signal(s) : FID1A.ch
 Acq On : 06 Dec 2018 10:32
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 07 00:48:31 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112818.M
 Quant Title :
 QLast Update : Thu Nov 29 04:54:55 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.715	5205621	47.767 ug/ml
Target Compounds			
2) N-DECANE	4.411	6593258	56.422 ug/ml
3) N-DODECANE	6.439	6628265	55.722 ug/ml
4) N-TETRADECANE	8.241	6616020	54.455 ug/ml
5) N-HEXADECANE	9.848	6575953	53.029 ug/ml
6) N-OCTADECANE	11.294	6469879	51.469 ug/ml
7) N-EICOSANE	12.607	6264651	49.922 ug/ml
8) N-DOCOSANE	13.811	6051594	48.391 ug/ml
10) N-TETRACOSANE	14.919	5839844	46.985 ug/ml
11) N-HEXACOSANE	15.946	5679787	45.714 ug/ml
12) N-OCTACOSANE	16.902	5579762	44.773 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120618\
Data File : FG003171.D
Signal(s) : FID1A.ch
Acq On : 06 Dec 2018 10:32
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 07 00:48:31 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112818.M
Quant Title :
QLast Update : Thu Nov 29 04:54:55 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

