

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120623\
 Data File : FG013243.D
 Signal(s) : FID1A.ch
 Acq On : 06 Dec 2023 14:20
 Operator : YP\AJ
 Sample : PB157588BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB157588BS

Integration File: autoint1.e
 Quant Time: Dec 07 00:06:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120423.M
 Quant Title :
 QLast Update : Mon Dec 04 11:42:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.081	1857023	17.298 ug/ml
Target Compounds			
2) N-DECANE	4.588	2191955	17.341 ug/ml
3) N-DODECANE	6.764	2201899	17.531 ug/ml
4) N-TETRADECANE	8.598	2146688	17.783 ug/ml
5) N-HEXADECANE	10.212	2158868	17.762 ug/ml
6) N-OCTADECANE	11.661	2118565	17.628 ug/ml
7) N-EICOSANE	12.975	2062990	17.458 ug/ml
8) N-DOCOSANE	14.178	1975790	17.419 ug/ml
10) N-TETRACOSANE	15.284	2022598	17.367 ug/ml
11) N-HEXACOSANE	16.308	1937682	17.399 ug/ml
12) N-OCTACOSANE	17.260	1954868	17.714 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120623\
Data File : FG013243.D
Signal(s) : FID1A.ch
Acq On : 06 Dec 2023 14:20
Operator : YP\AJ
Sample : PB157588BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB157588BS

Integration File: autoint1.e
Quant Time: Dec 07 00:06:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120423.M
Quant Title :
QLast Update : Mon Dec 04 11:42:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

