

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120720\
 Data File : FG007431.D
 Signal(s) : FID1A.ch
 Acq On : 07 Dec 2020 13:35
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 07 14:20:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120220.M
 Quant Title :
 QLast Update : Thu Dec 03 08:50:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.735	5950978	46.410 ug/ml
Target Compounds			
2) N-DECANE	4.436	7548461	50.437 ug/ml
3) N-DODECANE	6.464	7532187	49.987 ug/ml
4) N-TETRADECANE	8.266	7471182	49.306 ug/ml
5) N-HEXADECANE	9.872	7319864	48.397 ug/ml
6) N-OCTADECANE	11.319	7317625	47.761 ug/ml
7) N-EICOSANE	12.632	7184096	47.196 ug/ml
8) N-DOCOSANE	13.834	6944444	46.702 ug/ml
10) N-TETRACOSANE	14.942	6811443	46.041 ug/ml
11) N-HEXACOSANE	15.964	6526603	45.347 ug/ml
12) N-OCTACOSANE	16.915	6216460	44.773 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG120720\
Data File : FG007431.D
Signal(s) : FID1A.ch
Acq On : 07 Dec 2020 13:35
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 07 14:20:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG120220.M
Quant Title :
QLast Update : Thu Dec 03 08:50:10 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

