

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG12092
Data File : FG014936.D
Signal(s) : FID1A.ch
Acq On : 09 Dec 2024 09:28
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/10/2024
Supervised By :Ankita Jodhani 12/10/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112024.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.050	2.015	2.135	BB	568946	6135115	82.08%	5.120%
2	4.567	4.525	4.627	BB	728281	7347348	98.30%	6.132%
3	6.740	6.701	6.794	BB	735760	7452272	99.70%	6.220%
4	8.572	8.524	8.624	BB	692311	7452156	99.70%	6.220%
5	10.184	10.136	10.239	BB	675013	7413040	99.17%	6.187%
6	11.632	11.583	11.685	BB	644823	7474722	100.00%	6.239%
7	12.944	12.895	12.998	BB	614638	7470168	99.94%	6.235%
8	14.145	14.106	14.205	VB	565858	7148572	95.64%	5.966%
9	15.046	15.000	15.076	BV	454302	6360424	85.09%	5.309%
10	15.251	15.165	15.303	BB	544229	7024528	93.98%	5.863%
11	16.273	16.199	16.316	BB	497526	6883332	92.09%	5.745%
12	17.223	17.134	17.283	BB	505355	6813937	91.16%	5.687%
13	18.112	18.052	18.167	BB	503083	6997720	93.62%	5.840%
14	18.945	18.871	18.982	BV	470089	6885322	92.11%	5.747%
15	19.729	19.674	19.764	PV	444987	6377598	85.32%	5.323%
16	20.468	20.423	20.521	PV	372384	5675393	75.93%	4.737%
17	21.217	21.139	21.267	BV	272719	4803070	64.26%	4.009%
18	22.155	22.084	22.234	PV	161611	4100271	54.86%	3.422%
Sum of corrected areas:						119814987		

FG112024.M Tue Dec 10 01:56:40 2024