

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG121021\
 Data File : FG009223.D
 Signal(s) : FID1A.ch
 Acq On : 11 Dec 2021 04:56
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Dec 11 05:48:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112321.M
 Quant Title :
 QLast Update : Wed Nov 24 00:59:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.746	5270360	45.191 ug/ml
Target Compounds			
2) N-DECANE	4.431	7190601	48.545 ug/ml
3) N-DODECANE	6.460	7094355	47.926 ug/ml
4) N-TETRADECANE	8.264	6894292	47.649 ug/ml
5) N-HEXADECANE	9.872	6654244	47.061 ug/ml
6) N-OCTADECANE	11.322	6687501	46.117 ug/ml
7) N-EICOSANE	12.638	6544525	45.746 ug/ml
8) N-DOCOSANE	13.842	6283678	45.797 ug/ml
10) N-TETRACOSANE	14.951	6062877	46.068 ug/ml
11) N-HEXACOSANE	15.977	5736640	46.823 ug/ml
12) N-OCTACOSANE	16.932	5477254	48.039 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG121021\
Data File : FG009223.D
Signal(s) : FID1A.ch
Acq On : 11 Dec 2021 04:56
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Dec 11 05:48:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112321.M
Quant Title :
QLast Update : Wed Nov 24 00:59:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

