

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/12/2024
Supervised By :Ankita Jodhani 12/12/2024

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG12112
Data File : FG014970.D
Signal(s) : FID1A.ch
Acq On : 11 Dec 2024 14:46
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112024.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.571	4.527	4.629	BB	636060	6534952	95.79%	9.128%
2	6.745	6.698	6.788	BB	676897	6663083	97.66%	9.307%
3	8.578	8.518	8.644	BB	647491	6744858	98.86%	9.422%
4	10.191	10.141	10.236	BB	603594	6735029	98.72%	9.408%
5	11.638	11.530	11.686	BB	588291	6822521	100.00%	9.530%
6	12.952	12.869	13.001	BB	582484	6818540	99.94%	9.525%
7	14.154	14.114	14.213	VB	520562	6518953	95.55%	9.106%
8	15.055	15.000	15.088	BV	421352	5723969	83.90%	7.996%
9	15.261	15.110	15.309	BB	493348	6373033	93.41%	8.902%
10	16.283	16.111	16.326	BV	485396	6263121	91.80%	8.749%
11	17.234	17.122	17.293	VV	454304	6390965	93.67%	8.927%
Sum of corrected areas:						71589024		

FG112024.M Wed Dec 11 23:59:44 2024