

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG121621\
 Data File : FG009307.D
 Signal(s) : FID1A.ch
 Acq On : 16 Dec 2021 16:29
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/17/2021
 Supervised By :Ankita Jodhani 12/17/2021

Integration File: autoint1.e
 Quant Time: Dec 17 00:35:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121521.M
 Quant Title :
 QLast Update : Thu Dec 16 05:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.525	6776881	49.065 ug/ml
Target Compounds			
1) N-OCTANE	2.166	8391116	55.710 ug/ml
2) N-DECANE	4.243	8684920	56.506 ug/ml
3) N-DODECANE	6.266	8729471	55.458 ug/ml
4) N-TETRADECANE	8.064	8698237	54.112 ug/ml
5) N-HEXADECANE	9.667	8512444	53.115 ug/ml
6) N-OCTADECANE	11.112	8486290	52.146 ug/ml
7) N-EICOSANE	12.423	8284976	51.031 ug/ml
8) N-DOCOSANE	13.624	7930751	49.808 ug/ml
10) N-TETRACOSANE	14.729	7682769	48.674 ug/ml
11) N-HEXACOSANE	15.752	7343553	47.632 ug/ml
12) N-OCTACOSANE	16.703	7082423	47.331 ug/ml
13) N-TRIACONTANE	17.592	6862994	47.882 ug/ml
14) N-DOTRIACONTANE	18.425	6587790	49.164 ug/ml
15) N-TETRATRIACONTANE	19.209	6381246	50.894 ug/mlm
16) N-HEXATRIACONTANE	19.950	6103257	52.819 ug/ml
17) N-OCTATRIACONTANE	20.666	5888837	53.502 ug/mlm
18) N-TETRACONTANE	21.532	6010543	54.631 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG121621\
Data File : FG009307.D
Signal(s) : FID1A.ch
Acq On : 16 Dec 2021 16:29
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 12/17/2021
Supervised By : Ankita Jodhani 12/17/2021

Integration File: autoint1.e
Quant Time: Dec 17 00:35:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121521.M
Quant Title :
QLast Update : Thu Dec 16 05:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

