

rteres

Instrument :
FID_G

LabSampleId :
50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/17/2021
Supervised By :Ankita Jodhani 12/17/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG12162
Data File : FG009307.D
Signal(s) : FID1A.ch
Acq On : 16 Dec 2021 16:29
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121521.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.166	2.135	2.250	BB	1152423	8391116	96.12%	6.242%
2	4.243	4.215	4.329	BB	938962	8684920	99.49%	6.460%
3	6.266	6.227	6.341	BB	856730	8729471	100.00%	6.494%
4	8.064	8.026	8.134	BB	808471	8698237	99.64%	6.470%
5	9.667	9.629	9.725	BB	763903	8512444	97.51%	6.332%
6	11.112	11.069	11.166	BB	718547	8486290	97.21%	6.313%
7	12.423	12.377	12.480	BB	690989	8284976	94.91%	6.163%
8	13.624	13.585	13.679	VB	622283	7930751	90.85%	5.899%
9	14.525	14.466	14.593	BB	474004	6776881	77.63%	5.041%
10	14.729	14.663	14.780	BB	591985	7682769	88.01%	5.715%
11	15.752	15.698	15.803	BB	544106	7343553	84.12%	5.463%
12	16.703	16.632	16.750	BB	535232	7082423	81.13%	5.268%
13	17.592	17.526	17.630	BB	486173	6862994	78.62%	5.105%
14	18.425	18.366	18.463	BB	456394	6587790	75.47%	4.900%
15	19.210	18.887	19.247	BB	446121	6356305	72.81%	4.728%
16	19.950	19.884	19.988	BV	416488	6103257	69.92%	4.540%
17	20.666	20.609	20.710	M	346312	5906709	67.66%	4.394%
18	21.532	21.463	21.581	VBA	256589	6010543	68.85%	4.471%
Sum of corrected areas:						134431429		

FG121521.M Fri Dec 17 11:01:28 2021