

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122018\
 Data File : FB019137.D
 Signal(s) : FID2B.CH
 Acq On : 20 Dec 2018 8:26
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Dec 21 00:18:54 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
 Quant Title :
 QLast Update : Fri Dec 07 17:57:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.103	186300	21.314 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.957	445148	29.118 ng/ml
2) t 2,2,4-Trimethylpentane	7.721	460097	29.865 ng/ml
3) t n-Heptane	8.059	160403	10.792 ng/ml
4) t Benzene	8.198	196475	10.908 ng/ml
6) t Toluene	10.953	450021	29.648 ng/ml
7) t Ethylbenzene	13.404	149204	9.604 ng/ml
8) t m-Xylene	13.536	316308	19.415 ng/ml
9) t O-Xylene	14.268	326646	19.319 ng/ml
10) t 1,2,4-Trimethylbenzene	16.551	400411	19.785 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122018\
Data File : FB019137.D
Signal(s) : FID2B.CH
Acq On : 20 Dec 2018 8:26
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 21 00:18:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
Quant Title :
QLast Update : Fri Dec 07 17:57:38 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.957 min
Delta R.T.: -0.004 min
Response: 445148
Conc: 29.12 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.721 min
Delta R.T.: -0.002 min
Response: 460097
Conc: 29.86 ng/ml




#3 n-Heptane

R.T.: 8.059 min
Delta R.T.: -0.003 min
Response: 160403
Conc: 10.79 ng/ml




#4 Benzene

R.T.: 8.198 min
Delta R.T.: -0.003 min
Response: 196475
Conc: 10.91 ng/ml




#5 AAA-TFT

R.T.: 9.103 min
 Delta R.T.: -0.003 min
 Response: 186300
 Conc: 21.31 ng/ml




#6 Toluene

R.T.: 10.953 min
 Delta R.T.: -0.002 min
 Response: 450021
 Conc: 29.65 ng/ml




#7 Ethylbenzene

R.T.: 13.404 min
 Delta R.T.: -0.002 min
 Response: 149204
 Conc: 9.60 ng/ml




#8 m-Xylene

R.T.: 13.536 min
 Delta R.T.: -0.001 min
 Response: 316308
 Conc: 19.42 ng/ml



#9 O-Xylene

R.T.: 14.268 min
Delta R.T.: -0.002 min
Response: 326646
Conc: 19.32 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.551 min
Delta R.T.: -0.002 min
Response: 400411
Conc: 19.79 ng/ml