

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG122018\
 Data File : FG003312.D
 Signal(s) : FID1A.ch
 Acq On : 20 Dec 2018 13:44
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 20 23:50:48 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121318.M
 Quant Title :
 QLast Update : Thu Dec 13 00:59:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.708	5712876	46.061 ug/ml
Target Compounds			
2) N-DECANE	4.410	7183663	53.542 ug/ml
3) N-DODECANE	6.438	7286441	53.557 ug/ml
4) N-TETRADECANE	8.238	7295843	52.866 ug/ml
5) N-HEXADECANE	9.843	7247996	51.733 ug/ml
6) N-OCTADECANE	11.289	7124859	50.289 ug/ml
7) N-EICOSANE	12.603	6874435	48.661 ug/ml
8) N-DOCOSANE	13.805	6591948	46.868 ug/ml
10) N-TETRACOSANE	14.913	6302217	45.025 ug/ml
11) N-HEXACOSANE	15.939	6097185	43.745 ug/ml
12) N-OCTACOSANE	16.893	6013708	43.035 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG122018\
Data File : FG003312.D
Signal(s) : FID1A.ch
Acq On : 20 Dec 2018 13:44
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 20 23:50:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121318.M
Quant Title :
QLast Update : Thu Dec 13 00:59:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

