

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG122221\
 Data File : FG009330.D
 Signal(s) : FID1A.ch
 Acq On : 22 Dec 2021 15:07
 Operator : AJ\MA
 Sample : PB141630BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB141630BS

Integration File: autoint1.e
 Quant Time: Dec 22 23:54:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121521.M
 Quant Title :
 QLast Update : Thu Dec 16 05:01:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.517	2697529	19.530 ug/ml
Target Compounds			
2) N-DECANE	4.237	2988849	19.446 ug/ml
3) N-DODECANE	6.257	3094485	19.659 ug/ml
4) N-TETRADECANE	8.055	3156736	19.638 ug/ml
5) N-HEXADECANE	9.658	2964520	18.498 ug/ml
6) N-OCTADECANE	11.102	3196278	19.640 ug/ml
7) N-EICOSANE	12.415	3132420	19.294 ug/ml
8) N-DOCOSANE	13.615	3023634	18.989 ug/ml
10) N-TETRACOSANE	14.721	2957803	18.739 ug/ml
11) N-HEXACOSANE	15.743	2802866	18.180 ug/ml
12) N-OCTACOSANE	16.696	2682356	17.926 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG122221\
Data File : FG009330.D
Signal(s) : FID1A.ch
Acq On : 22 Dec 2021 15:07
Operator : AJ\MA
Sample : PB141630BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB141630BS

Integration File: autoint1.e
Quant Time: Dec 22 23:54:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121521.M
Quant Title :
QLast Update : Thu Dec 16 05:01:34 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

