

Method Path : P:\HPCHEM1\FID_F\Method\
 Method File : FG022818.M
 Title :
 Last Update : Wed Feb 28 18:10:38 2018
 Response Via : Initial Calibration

Calibration Files

50 =FG000126.D 100 =FG000125.D 20 =FG000127.D
 10 =FG000128.D 5 =FG000129.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	1.379	1.336	1.394	1.460	1.449	1.404	E5	3.67
2)	N-DECANE	1.392	1.341	1.406	1.475	1.464	1.416	E5	3.88
3)	N-DODECANE	1.375	1.317	1.390	1.462	1.454	1.400	E5	4.27
4)	N-TETRADECANE	1.348	1.289	1.363	1.437	1.437	1.375	E5	4.58
5)	N-HEXADECANE	1.301	1.248	1.311	1.380	1.378	1.323	E5	4.23
6)	N-OCTADECANE	1.287	1.235	1.315	1.418	1.368	1.324	E5	5.36
7)	N-EICOSANE	1.246	1.202	1.262	1.339	1.378	1.285	E5	5.57
8)	N-DOCOSANE	1.215	1.183	1.214	1.266	1.259	1.227	E5	2.82
9) S	TETRACOSANE-d50 (S...	1.099	1.067	1.094	1.156	1.142	1.112	E5	3.29
10)	N-TETRACOSANE	1.209	1.181	1.209	1.259	1.260	1.224	E5	2.84
11)	N-HEXACOSANE	1.195	1.172	1.192	1.242	1.233	1.207	E5	2.48
12)	N-OCTACOSANE	1.173	1.156	1.171	1.218	1.215	1.187	E5	2.37
13)	N-TRIACONTANE	1.132	1.122	1.136	1.191	1.186	1.153	E5	2.83
14)	N-DOTRIACONTANE	0.963	0.957	0.990	1.043	1.043	0.999	E5	4.20
15)	N-TETRATRIACONTANE	6.289	6.213	6.829	7.107	6.819	6.651	E4	5.78
16)	N-HEXATRIACONTANE	3.283	3.240	3.857	3.833	3.488	3.540	E4	8.30
17)	N-OCTATRIACONTANE	2.037	1.980	2.606	2.530	2.255	2.282	E4	12.36
18)	N-TETRACONTANE	1.477	1.428	1.987	1.792	1.528	1.642	E4	14.52

(#) = Out of Range