

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG121319\
Data File : FG006355.D
Signal(s) : FID1A.ch
Acq On : 13 Dec 2019 16:14
Operator : DD\AJ
Sample : RT MARKER
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 13 16:44:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112619.M
Quant Title :
QLast Update : Wed Nov 27 02:46:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
Target Compounds				
3)	N-DODECANE	6.401	2090626	21.439 ug/ml
12)	N-OCTACOSANE	16.635	2194198	23.864 ug/ml
13)	N-TRIACONTANE	17.508	399330	4.077 ug/ml
14)	N-DOTRIACONTANE	18.328	421968	4.592 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG121319\
Data File : FG006355.D
Signal(s) : FID1A.ch
Acq On : 13 Dec 2019 16:14
Operator : DD\AJ
Sample : RT MARKER
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 13 16:44:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG112619.M
Quant Title :
QLast Update : Wed Nov 27 02:46:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE


R.T.: 6.401 min
Delta R.T.: 0.028 min
Response: 2090626
Conc: 21.44 ug/ml

Instrument :
FID_G
ClientSampleId :




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.477 min
Response: 0
Conc: N.D.




#12 N-OCTACOSANE

R.T.: 16.635 min
Delta R.T.: 0.020 min
Response: 2194198
Conc: 23.86 ug/ml



#13 N-TRIACONTANE

R.T.: 17.508 min
Delta R.T.: 0.021 min
Response: 399330
Conc: 4.08 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.328 min
Delta R.T.: 0.022 min
Response: 421968
Conc: 4.59 ug/ml

Instrument :
FID_G
ClientSampleId :