

Data Path : P:\HPCHEM1\FID_G\Data\FG030918\
 Data File : FG000323.D
 Signal(s) : FID1A.ch
 Acq On : 09 Mar 2018 12:40
 Operator : UA\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 09 14:13:47 2018
 Quant Method : P:\HPCHEM1\FID_F\Method\FG022818.M
 Quant Title :
 QLast Update : Wed Feb 28 18:10:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.573	5583302	50.232 ug/ml
Target Compounds			
2) N-DECANE	4.311	6839863	48.316 ug/ml
3) N-DODECANE	6.330	6643604	47.463 ug/ml
4) N-TETRADECANE	8.124	6493165	47.230 ug/ml
5) N-HEXADECANE	9.724	6326696	47.805 ug/ml
6) N-OCTADECANE	11.166	6283933	47.446 ug/ml
7) N-EICOSANE	12.475	6189707	48.159 ug/ml
8) N-DOCOSANE	13.673	6228868	50.746 ug/ml
10) N-TETRACOSANE	14.777	6280471	51.322 ug/ml
11) N-HEXACOSANE	15.798	6296334	52.178 ug/ml
12) N-OCTACOSANE	16.748	6296457	53.064 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : P:\HPCHEM1\FID_G\Data\FG030918\
Data File : FG000323.D
Signal(s) : FID1A.ch
Acq On : 09 Mar 2018 12:40
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 09 14:13:47 2018
Quant Method : P:\HPCHEM1\FID_F\Method\FG022818.M
Quant Title :
QLast Update : Wed Feb 28 18:10:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

