

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031219\
 Data File : FG004153.D
 Signal(s) : FID1A.ch
 Acq On : 13 Mar 2019 01:44
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Sohil
 3/13/2019 11:57:55 AM

Integration File: autoint1.e
 Quant Time: Mar 13 02:07:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
 Quant Title :
 QLast Update : Tue Mar 05 03:05:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.438	5480499	43.593 ug/ml
Target Compounds			
1) N-OCTANE	2.089	6457058	49.129 ug/ml
2) N-DECANE	4.170	7060419	52.809 ug/ml
3) N-DODECANE	6.193	7098630	52.362 ug/ml
4) N-TETRADECANE	7.987	7058795	50.621 ug/ml
5) N-HEXADECANE	9.588	6964222	48.934 ug/ml
6) N-OCTADECANE	11.029	6809590	47.462 ug/ml
7) N-EICOSANE	12.340	6587092	46.110 ug/ml
8) N-DOCOSANE	13.539	6350221	44.739 ug/ml
10) N-TETRACOSANE	14.642	6160230	43.681 ug/ml
11) N-HEXACOSANE	15.663	6025708	43.265 ug/ml
12) N-OCTACOSANE	16.616	6052254	44.206 ug/ml
13) N-TRIACONTANE	17.504	6239072	46.718 ug/ml
14) N-DOTRIACONTANE	18.337	6264821	49.512 ug/ml
15) N-TETRATRIACONTANE	19.123	6093359	51.420 ug/ml
16) N-HEXATRIACONTANE	19.866	5953800	53.780 ug/mlm
17) N-OCTATRIACONTANE	20.580	5510730	54.591 ug/ml
18) N-TETRACONTANE	21.413	5696778	53.711 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031219\
Data File : FG004153.D
Signal(s) : FID1A.ch
Acq On : 13 Mar 2019 01:44
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampled :
50 PPM TRPH STD

Manual Integrations
APPROVED

Sohil
3/13/2019 11:57:55 AM

Integration File: autoint1.e
Quant Time: Mar 13 02:07:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
Quant Title :
QLast Update : Tue Mar 05 03:05:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

