

Data Path : \\74.0.250.170\pestpcbsrv\HPCHEM1\FID_G\Data\FG031418\
 Data File : FG000405.D
 Signal(s) : FID1A.ch
 Acq On : 14 Mar 2018 17:14
 Operator : UA\AJ
 Sample : J1908-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 15 08:29:28 2018
 Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\FID_F\Method\FG022818.M
 Quant Title :
 QLast Update : Wed Feb 28 18:10:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.572	3172411	28.541 ug/ml
Target Compounds			
4) N-TETRADECANE	8.176f	578419	4.207 ug/ml
5) N-HEXADECANE	9.722	1024710	7.743 ug/ml
7) N-EICOSANE	12.477	540864	4.208 ug/ml
8) N-DOCOSANE	13.685	2462344	20.060 ug/ml
10) N-TETRACOSANE	14.757	11005936	89.937 ug/ml
11) N-HEXACOSANE	15.797	602330	4.991 ug/ml
12) N-OCTACOSANE	16.749	3452961	29.100 ug/ml
13) N-TRIACONTANE	17.633	607694	5.268 ug/ml
14) N-DOTRIACONTANE	18.463	192170	1.924 ug/ml
15) N-TETRATRIACONTANE	19.282	366837	5.515 ug/ml
16) N-HEXATRIACONTANE	19.953	868733	24.539 ug/ml
18) N-TETRACONTANE	21.605	803744	48.940 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : \\74.0.250.170\pestpcbsrv\HPCHEM1\FID_G\Data\FG031418\
Data File : FG000405.D
Signal(s) : FID1A.ch
Acq On : 14 Mar 2018 17:14
Operator : UA\AJ
Sample : J1908-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 15 08:29:28 2018
Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\FID_F\Method\FG022818.M
Quant Title :
QLast Update : Wed Feb 28 18:10:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

