

Data Path : P:\HPCHEM1\FID_G\Data\FG031418\
 Data File : FG000422.D
 Signal(s) : FID1A.ch
 Acq On : 15 Mar 2018 01:22
 Operator : UA\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Mar 15 08:32:12 2018
 Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\FID_F\Method\FG022818.M
 Quant Title :
 QLast Update : Wed Feb 28 18:10:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.573	5060246	45.526 ug/mlm
Target Compounds			
2) N-DECANE	4.306	6811751	48.118 ug/ml
3) N-DODECANE	6.329	6650387	47.511 ug/ml
4) N-TETRADECANE	8.124	6458361	46.976 ug/ml
5) N-HEXADECANE	9.724	6195409	46.813 ug/ml
6) N-OCTADECANE	11.166	6059394	45.750 ug/ml
7) N-EICOSANE	12.475	5807183	45.183 ug/ml
8) N-DOCOSANE	13.673	5680994	46.282 ug/ml
10) N-TETRACOSANE	14.777	5658453	46.239 ug/ml
11) N-HEXACOSANE	15.798	5645029	46.780 ug/ml
12) N-OCTACOSANE	16.749	5664165	47.736 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : P:\HPCHEM1\FID_G\Data\FG031418\
Data File : FG000422.D
Signal(s) : FID1A.ch
Acq On : 15 Mar 2018 01:22
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Mar 15 08:32:12 2018
Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\FID_F\Method\FG022818.M
Quant Title :
QLast Update : Wed Feb 28 18:10:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

