

Method Path : P:\HPCHEM1\FID_F\Method\
 Method File : FF040918.M
 Title :
 Last Update : Mon Apr 09 14:25:39 2018
 Response Via : Initial Calibration

Calibration Files

50 =FF000368.D 100 =FF000367.D 20 =FF000369.D
 10 =FF000370.D 5 =FF000371.D

Compound		50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.010	0.825	0.947	0.771	0.936	0.898 E5	10.84
2)	N-DECANE	1.116	0.928	1.034	0.844	1.012	0.987 E5	10.56
3)	N-DODECANE	1.255	1.092	1.129	0.941	1.107	1.105 E5	10.13
4)	N-TETRADECANE	1.404	1.270	1.246	1.094	1.239	1.251 E5	8.82
5)	N-HEXADECANE	1.509	1.391	1.343	1.256	1.384	1.377 E5	6.66
6)	N-OCTADECANE	1.581	1.469	1.444	1.486	1.693	1.535 E5	6.69
7)	N-EICOSANE	1.569	1.487	1.414	1.477	1.589	1.507 E5	4.75
8)	N-DOCOSANE	1.532	1.494	1.363	1.446	1.463	1.460 E5	4.31
9) S	TETRACOSANE-d50 (S...	1.320	1.332	1.188	1.308	1.305	1.291 E5	4.52
10)	N-TETRACOSANE	1.479	1.508	1.342	1.473	1.463	1.453 E5	4.44
11)	N-HEXACOSANE	1.366	1.504	1.316	1.523	1.465	1.435 E5	6.26
12)	N-OCTACOSANE	1.189	1.417	1.259	1.611	1.503	1.396 E5	12.40
13)	N-TRIACONTANE	1.000	1.231	1.099	1.529	1.364	1.244 E5	16.88
14)	N-DOTRIACONTANE	0.862	1.012	0.905	1.223	1.093	1.019 E5	14.30
15)	N-TETRATRIACONTANE	8.434	9.275	7.835	9.976	9.481	9.000 E4	9.52
16)	N-HEXATRIACONTANE	8.670	9.170	7.514	9.002	8.811	8.633 E4	7.57
17)	N-OCTATRIACONTANE	8.457	8.947	7.141	8.706	8.668	8.384 E4	8.55
18)	N-TETRACONTANE	8.896	9.371	7.920	9.146	9.091	8.885 E4	6.36

(#) = Out of Range