

Method Path : P:\HPCHEM1\FID_G\Method\
 Method File : FG040918.M
 Title :
 Last Update : Mon Apr 09 14:24:49 2018
 Response Via : Initial Calibration

Calibration Files

50 =FG000787.D 100 =FG000786.D 20 =FG000788.D
 10 =FG000789.D 5 =FG000790.D

	Compound	50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.355	1.478	1.621	1.543	1.678	1.535 E5	8.20
2)	N-DECANE	1.501	1.629	1.730	1.659	1.798	1.664 E5	6.73
3)	N-DODECANE	1.701	1.834	1.844	1.801	1.916	1.819 E5	4.31
4)	N-TETRADECANE	1.956	2.022	1.997	2.044	2.100	2.024 E5	2.65
5)	N-HEXADECANE	2.139	2.082	2.105	2.302	2.310	2.188 E5	5.02
6)	N-OCTADECANE	2.241	2.012	2.148	2.518	2.556	2.295 E5	10.27
7)	N-EICOSANE	2.203	1.820	1.983	2.373	2.289	2.133 E5	10.67
8)	N-DOCOSANE	2.094	1.617	1.774	2.141	1.980	1.921 E5	11.51
9) S	TETRACOSANE-d50 (S...	1.747	1.328	1.485	1.819	1.734	1.623 E5	12.78
10)	N-TETRACOSANE	1.890	1.447	1.586	1.909	1.756	1.718 E5	11.58
11)	N-HEXACOSANE	1.639	1.334	1.445	1.690	1.566	1.535 E5	9.46
12)	N-OCTACOSANE	1.474	1.304	1.430	1.630	1.625	1.493 E5	9.23
13)	N-TRIACONTANE	1.421	1.334	1.505	1.681	1.821	1.552 E5	12.73
14)	N-DOTRIACONTANE	1.331	1.303	1.438	1.569	1.720	1.472 E5	11.80
15)	N-TETRATRIACONTANE	1.354	1.348	1.391	1.478	1.553	1.425 E5	6.21
16)	N-HEXATRIACONTANE	1.407	1.395	1.414	1.488	1.477	1.436 E5	3.01
17)	N-OCTATRIACONTANE	1.377	1.351	1.403	1.477	1.468	1.415 E5	3.91
18)	N-TETRACONTANE	1.435	1.400	1.467	1.536	1.521	1.472 E5	3.89

(#) = Out of Range