

Method Path : P:\HPCHEM1\FID_G\Method\
 Method File : FG042318.M
 Title :
 Last Update : Mon Apr 23 17:37:23 2018
 Response Via : Initial Calibration

Calibration Files

50 =FG000804.D 100 =FG000803.D 20 =FG000805.D
 10 =FG000806.D 5 =FG000807.D

	Compound	50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.605	1.598	1.636	1.667	1.620	1.625 E5	1.71
2)	N-DECANE	1.623	1.617	1.647	1.665	1.598	1.630 E5	1.61
3)	N-DODECANE	1.630	1.628	1.655	1.672	1.590	1.635 E5	1.90
4)	N-TETRADECANE	1.650	1.643	1.663	1.686	1.608	1.650 E5	1.75
5)	N-HEXADECANE	1.657	1.649	1.693	1.713	1.654	1.673 E5	1.69
6)	N-OCTADECANE	1.656	1.648	1.696	1.732	1.674	1.681 E5	2.01
7)	N-EICOSANE	1.632	1.626	1.674	1.714	1.666	1.662 E5	2.13
8)	N-DOCOSANE	1.614	1.613	1.657	1.701	1.663	1.650 E5	2.23
9) S	TETRACOSANE-d50 (S...	1.413	1.409	1.451	1.490	1.457	1.444 E5	2.32
10)	N-TETRACOSANE	1.591	1.588	1.632	1.676	1.636	1.625 E5	2.25
11)	N-HEXACOSANE	1.580	1.576	1.620	1.664	1.621	1.612 E5	2.22
12)	N-OCTACOSANE	1.581	1.562	1.613	1.664	1.629	1.610 E5	2.50
13)	N-TRIACONTANE	1.593	1.548	1.611	1.674	1.648	1.615 E5	3.03
14)	N-DOTRIACONTANE	1.569	1.507	1.579	1.642	1.616	1.583 E5	3.24
15)	N-TETRATRIACONTANE	1.550	1.498	1.569	1.614	1.575	1.561 E5	2.73
16)	N-HEXATRIACONTANE	1.473	1.440	1.507	1.536	1.491	1.490 E5	2.42
17)	N-OCTATRIACONTANE	1.323	1.302	1.355	1.371	1.330	1.336 E5	2.01
18)	N-TETRACONTANE	1.296	1.286	1.319	1.323	1.271	1.299 E5	1.70

(#) = Out of Range