

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010421\
 Data File : PS013649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2021 16:27
 Operator : DD\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Ankita
 1/5/2021 1:52:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 01:33:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:31:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	6.913	181.1E6	77494876	192.986	196.947
Target Compounds							
1) T	Dalapon	3.029	2.621	180.9E6	91302574	168.809	179.809
2) T	3,5-DICHL...	6.428	6.069	234.5E6	113.9E6	179.891	183.711
3) T	4-Nitroph...	7.052	6.537	75181679	40310935	159.711	158.714
5) T	DICAMBA	7.398	7.077	686.8E6	314.2E6	180.058	184.237
6) T	MCPD	7.560	7.165	43535490	24287433	15.218	17.420
7) T	MCPA	7.709	7.379	66341205	34841418	16.257	17.596
8) T	DICHLORPROP	8.089	7.703	175.6E6	81586796	185.488	187.061
9) T	2,4-D	8.329	8.009	162.7E6	98567088	152.416	180.368
10) T	Pentachlo...	8.627	8.424	2777.2E6	1192.2E6	198.233	208.699
11) T	2,4,5-TP ...	9.179	8.788	1033.6E6	454.3E6	188.058	192.701
12) T	2,4,5-T	9.472	9.177	1038.8E6	401.8E6	185.733	174.899
13) T	2,4-DB	10.033	9.689	145.9E6	49071006	149.918	161.437
14) T	DINOSEB	11.168	10.002	834.7E6	301.9E6	195.032	183.592
15) T	Picloram	11.028	11.018	974.6E6	568.7E6	131.837m	162.826
16) T	DCPA	11.433	10.926	1365.7E6	615.5E6	196.622	202.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010421\
Data File : PS013649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2021 16:27
Operator : DD\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDICC200

Manual Integrations
APPROVED

Ankita
1/5/2021 1:52:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 01:33:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:31:32 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

