

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010421\
 Data File : PS013652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2021 17:36
 Operator : DD\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 01:34:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:31:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.207	6.909	960.5E6	401.5E6	1023.625	1020.414
Target Compounds							
1) T	Dalapon	3.029	2.622	1007.6E6	472.4E6	940.061	930.412
2) T	3,5-DICHL...	6.424	6.064	1231.7E6	582.8E6	944.963	940.180
3) T	4-Nitroph...	7.041	6.521	447.0E6	236.6E6	949.658	931.659
5) T	DICAMBA	7.397	7.076	3634.1E6	1601.4E6	952.778	939.092
6) T	MCPD	7.566	7.172	283.4E6	134.7E6	99.058	96.627
7) T	MCPA	7.716	7.388	394.8E6	190.1E6	96.754	96.000
8) T	DICHLORPROP	8.088	7.700	909.4E6	415.8E6	960.624	953.364
9) T	2,4-D	8.319	7.997	1042.6E6	522.9E6	976.762	956.887
10) T	Pentachlo...	8.625	8.422	13071.6E6	5212.7E6	933.027	912.503
11) T	2,4,5-TP ...	9.174	8.783	5254.6E6	2229.8E6	956.074	945.743
12) T	2,4,5-T	9.461	9.163	5358.7E6	2196.1E6	958.148	955.947
13) T	2,4-DB	10.014	9.669	988.7E6	299.4E6	1015.772	984.984
14) T	DINOSEB	11.167	9.998	4100.0E6	1575.1E6	957.955	957.919
15) T	Picloram	10.999	10.991	7438.3E6	3417.0E6	1006.222	978.396
16) T	DCPA	11.433	10.925	6661.5E6	2882.6E6	959.043	945.977

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010421\
Data File : PS013652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2021 17:36
Operator : DD\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 01:34:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:31:32 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

