

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010422\
 Data File : PS017901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2022 13:00
 Operator : AJ\MA
 Sample : M5340-01MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BAYSHORE-FILL MS

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/05/2022
 Supervised By :mohammad ahmed 01/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 04 23:08:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
 Quant Title : 8080.M
 QLast Update : Tue Dec 14 13:29:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.123	6.711	814.5E6	322.9E6	521.099	541.755m
Target Compounds							
1) T	Dalapon	2.975	2.484	908.6E6	524.5E6	354.567	677.510 #
2) T	3,5-DICHL...	6.352	5.914	1066.8E6	405.8E6	477.477	447.564m
3) T	4-Nitroph...	6.957	6.345	243.4E6	108.5E6	359.889m	311.638m
5) T	DICAMBA	7.311	6.868	2897.8E6	1270.6E6	458.270	476.960
6) T	MCP	7.478	6.961	225.9E6	99226548	47.928	45.588
7) T	MCPA	7.624	7.171	289.1E6	150.0E6	43.719m	49.060
8) T	DICHLORPROP	8.000	7.485	718.9E6	319.2E6	460.098	468.421
9) T	2,4-D	8.231	7.778	864.1E6	396.0E6	488.798	487.101
10) T	Pentachlo...	8.531	8.193	11240.7E6	4555.4E6	520.949	531.694
11) T	2,4,5-TP ...	9.080	8.553	4354.0E6	1845.3E6	511.822	507.621m
12) T	2,4,5-T	9.365	8.932	4037.1E6	1670.5E6	464.074	488.508
13) T	2,4-DB	9.915	9.436	508.4E6	213.5E6	344.267	471.238 #
14) T	DINOSEB	11.059	9.755	3087.8E6	1125.7E6	451.568	435.727
15) T	Picloram	10.888	10.740	4162.7E6	2024.9E6	383.617	412.279
16) T	DCPA	11.328	10.682	4694.9E6	2346.4E6	432.288	523.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010422\
 Data File : PS017901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2022 13:00
 Operator : AJ\MA
 Sample : M5340-01MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BAYSHORE-FILL MS

Manual Integrations APPROVED

Reviewed By :Ankita Jodhani 01/05/2022
 Supervised By :mohammad ahmed 01/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 04 23:08:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
 Quant Title : 8080.M
 QLast Update : Tue Dec 14 13:29:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

