

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010423\
 Data File : PS021622.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2023 20:38
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 00:25:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.374	6.808	2104.6E6	1363.1E6	724.357	790.684
Target Compounds							
1) T	Dalapon	2.480	2.556	2242.3E6	1659.1E6	709.557	739.043
2) T	3,5-DICHL...	5.746	5.989	2829.3E6	1969.6E6	711.894	752.684
3) T	4-Nitroph...	6.215	6.424	1149.6E6	819.0E6	639.331	702.788
5) T	DICAMBA	6.516	6.971	8517.3E6	5840.5E6	740.168	764.742
6) T	MCPD	6.673	7.071	641.6E6	454.3E6	68.713	77.217
7) T	MCPA	6.799	7.284	899.5E6	631.6E6	70.218	71.426
8) T	DICHLORPROP	7.113	7.597	2242.5E6	1468.8E6	713.313	745.740
9) T	2,4-D	7.316	7.889	2683.1E6	1799.1E6	702.169	745.793
10) T	Pentachlo...	7.543	8.315	30643.5E6	19827.0E6	753.546	752.099
11) T	2,4,5-TP ...	8.072	8.677	13387.3E6	8478.7E6	724.003	757.131
12) T	2,4,5-T	8.332	9.053	13793.3E6	8148.3E6	731.061	751.782
13) T	2,4-DB	8.841	9.560	2381.9E6	1102.9E6	650.392	767.238
14) T	DINOSEB	9.894	9.887	11525.0E6	6002.9E6	723.353	758.983
15) T	Picloram	9.730	10.859	18567.7E6	10876.6E6	648.313	668.852
16) T	DCPA	10.161	10.814	18349.9E6	10171.2E6	720.521	749.634

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010423\
Data File : PS021622.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2023 20:38
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 00:25:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

