

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010524\
 Data File : PS025471.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2024 23:05
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/08/2024
 Supervised By :Ankita Jodhani 01/08/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 23:31:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122823.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 28 17:23:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.350	6.761	1655.9E6	540.8E6	738.982m	761.868
Target Compounds							
1) T	Dalapon	2.469	2.526	1926.0E6	712.0E6	622.382	619.459
2) T	3,5-DICHL...	5.725	5.953	2369.8E6	809.1E6	696.784m	723.450
3) T	4-Nitroph...	6.190	6.380	1055.3E6	340.5E6	681.208	685.385
5) T	DICAMBA	6.490	6.921	7065.7E6	2214.0E6	724.185	731.445
6) T	MCP	6.641	7.017	508.6E6	149.2E6	75.595	74.751
7) T	MCPA	6.763	7.223	699.4E6	227.6E6	73.747	72.401
8) T	DICHLORPROP	7.077	7.533	1827.3E6	558.3E6	710.304	721.740
9) T	2,4-D	7.276	7.816	2207.2E6	687.5E6	712.498	716.775
10) T	Pentachlo...	7.502	8.236	26871.3E6	8274.5E6	762.307	746.387
11) T	2,4,5-TP ...	8.023	8.594	10986.7E6	3185.6E6	722.928	736.758
12) T	2,4,5-T	8.280	8.964	10537.5E6	3055.3E6	702.562m	715.869
13) T	2,4-DB	8.785	9.467	1586.0E6	383.4E6	657.715	698.994
14) T	DINOSEB	9.834	9.793	8462.3E6	2279.3E6	736.217	717.660
15) T	Picloram	9.673	10.749	12625.0E6	3920.8E6	604.276	575.454
16) T	DCPA	10.099	10.712	13826.6E6	4158.8E6	727.576m	778.752

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010524\
Data File : PS025471.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2024 23:05
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/08/2024
Supervised By : Ankita Jodhani 01/08/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 23:31:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122823.M
Quant Title : 8080.M
QLast Update : Thu Dec 28 17:23:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

