

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010622\
 Data File : PS017920.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jan 2022 13:44
 Operator : AJ\MA
 Sample : N1014-02MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 72-11944MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/07/2022
 Supervised By : Ankita Jodhani 01/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 00:07:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.052	6.646	737.0E6	295.4E6	509.836	509.115
Target Compounds							
1) T	Dalapon	2.913	2.423	940.4E6	389.6E6	383.319	509.445 #
2) T	3,5-DICHL...	6.290	5.861	1049.8E6	371.7E6	488.003	420.127m
5) T	DICAMBA	7.237	6.800	2673.2E6	1143.5E6	428.100	434.931
6) T	MCP	7.406	6.892	235.1E6	94162760	50.364	43.778
7) T	MCPA	7.552	7.100	285.6E6	146.0E6	42.811	48.915
8) T	DICHLORPROP	7.925	7.411	715.0E6	327.9E6	454.552	491.319
9) T	2,4-D	8.153	7.700	908.3E6	389.8E6	498.802	529.615
10) T	Pentachlo...	8.451	8.113	10265.4E6	4022.7E6	485.914	477.625
11) T	2,4,5-TP ...	9.003	8.476	4283.7E6	1757.2E6	506.224	492.954
12) T	2,4,5-T	9.286	8.854	4197.2E6	1576.1E6	473.505	481.834
13) T	2,4-DB	9.837	9.356	544.3E6	201.0E6	329.255	454.377 #
14) T	DINOSEB	10.981	9.677	2685.9E6	1099.9E6	409.001	440.836
15) T	Picloram	10.805	10.654	5180.6E6	1919.2E6	434.038	410.433m
16) T	DCPA	11.251	10.603	5124.7E6	2193.3E6	474.302	496.735

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010622\
Data File : PS017920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jan 2022 13:44
Operator : AJ\MA
Sample : N1014-02MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
72-11944MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/07/2022
Supervised By : Ankita Jodhani 01/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 00:07:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

