

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010821\
 Data File : PS013716.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2021 00:48
 Operator : DD\AJ
 Sample : M1019-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 20410MSD

Manual Integrations
 APPROVED

Ankita
 1/11/2021 1:55:27 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 02:54:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:51:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	6.910	326.3E6	160.9E6	341.481	400.049
Target Compounds							
1) T	Dalapon	3.026	2.619	506.1E6	304.9E6	465.361m	586.363m#
2) T	3,5-DICHL...	6.424	6.064	655.0E6	281.0E6	495.315	445.953
5) T	DICAMBA	7.398	7.075	1911.1E6	863.2E6	496.258	502.635
6) T	MCP	7.564	7.168	153.7E6	71788657	53.664	50.801
7) T	MCPA	7.713	7.382	237.4E6	131.0E6	57.765	64.812m
8) T	DICHLORPROP	8.089	7.700	498.0E6	229.1E6	514.507	514.472
9) T	2,4-D	8.322	7.997	609.5E6	307.1E6	575.403	553.990
10) T	Pentachlo...	8.628	8.422	5662.1E6	2384.5E6	399.663	412.488
11) T	2,4,5-TP ...	9.175	8.781	3051.3E6	1905.1E6	547.967	797.518 #
12) T	2,4,5-T	9.463	9.163	3010.3E6	1257.0E6	532.357	549.074
13) T	2,4-DB	10.015	9.670	320.4E6	161.1E6	327.856m	529.710 #
14) T	DINOSEB	11.169	9.998	1443.7E6	547.2E6	327.225	327.041
15) T	Picloram	11.005	10.994	3025.1E6	1490.3E6	418.832	428.800
16) T	DCPA	11.435	10.925	3776.4E6	1694.0E6	534.504	547.334

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS010821\
Data File : PS013716.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2021 00:48
Operator : DD\AJ
Sample : M1019-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
20410MSD

Manual Integrations
APPROVED

Ankita
1/11/2021 1:55:27 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 02:54:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:51:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

