

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011023\
 Data File : PS021709.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jan 2023 21:54
 Operator : AR\AJ
 Sample : 01053-01MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-201MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 01:06:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010923.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 10 10:34:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.376	6.811	1216.0E6	723.3E6	436.850	402.592
Target Compounds							
2) T	3,5-DICHL...	5.748	5.992	1436.2E6	932.9E6	371.280	338.527
5) T	DICAMBA	6.517	6.973	4307.0E6	2924.1E6	352.253	343.900
6) T	MCPD	6.670	7.069	286.4E6	199.6E6	29.886	30.098
7) T	MCPA	6.794	7.280	397.5E6	315.2E6	30.772	34.754
8) T	DICHLORPROP	7.115	7.599	1090.2E6	713.5E6	338.019	334.400
9) T	2,4-D	7.318	7.891	1585.7E6	958.6E6	408.282	379.399
10) T	Pentachlo...	7.544	8.318	3389.2E6	1877.7E6	81.366	65.428
11) T	2,4,5-TP ...	8.074	8.680	5837.5E6	3684.9E6	316.807	309.902
12) T	2,4,5-T	8.334	9.057	6336.5E6	3717.8E6	328.730	332.839
13) T	2,4-DB	8.844	9.563	868.8E6	399.5E6	274.463	266.670
15) T	Picloram	9.734	10.863	9138.3E6	5639.7E6	322.140	318.475
16) T	DCPA	10.162	10.818	10215.0E6	5447.5E6	413.498	390.789

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011023\
Data File : PS021709.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jan 2023 21:54
Operator : AR\AJ
Sample : 01053-01MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-201MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 01:06:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010923.M
Quant Title : 8080.M
QLast Update : Tue Jan 10 10:34:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

