

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011025\
 Data File : PS028878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jan 2025 09:34
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 01:45:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122324.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 23 13:44:25 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.204	7.685	1768.5E6	763.6E6	781.819	678.400
Target Compounds							
1) T	Dalapon	2.617	2.668	2001.0E6	1247.7E6	710.417	633.825
2) T	3,5-DICHL...	6.380	6.648	2484.1E6	1069.1E6	753.859	644.976m
3) T	4-Nitroph...	7.003	7.214	1068.1E6	529.7E6	711.058	632.889
5) T	DICAMBA	7.390	7.882	7604.0E6	3667.6E6	771.143	668.067
6) T	MCPD	7.571	7.985	465.1E6	204.7E6	74.682	62.654
7) T	MCPA	7.720	8.228	629.2E6	284.6E6	71.783	61.426
8) T	DICHLORPROP	8.094	8.595	2000.6E6	912.0E6	754.022	661.001
9) T	2,4-D	8.324	8.921	2111.0E6	934.9E6	740.822m	647.841m
10) T	Pentachlo...	8.621	9.446	31086.9E6	15152.0E6	795.733	695.088
11) T	2,4,5-TP ...	9.197	9.822	12223.1E6	6160.3E6	778.082	685.736
12) T	2,4,5-T	9.488	10.240	12363.1E6	5825.2E6	759.909	668.853
13) T	2,4-DB	10.059	10.805	2235.1E6	607.0E6	716.023	625.103
14) T	DINOSEB	11.264	11.182	10100.8E6	3999.0E6	751.450	649.284
15) T	Picloram	11.074	12.267	19427.9E6	8226.1E6	724.575	640.714m
16) T	DCPA	11.559	12.221	18508.8E6	7590.2E6	769.565m	706.426m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011025\
Data File : PS028878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jan 2025 09:34
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 01:45:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122324.M
Quant Title : 8080.M
QLast Update : Mon Dec 23 13:44:25 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

