

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011122\
 Data File : PS017957.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2022 11:27
 Operator : AJ\MA
 Sample : N1052-01MS
 Misc : LOT#M03961
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BALL-1MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/12/2022
 Supervised By : Ankita Jodhani 01/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 15:29:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.051	6.645	875.3E6	304.7E6	605.540	525.175
Target Compounds							
1) T	Dalapon	2.909	2.417	507.1E6	357.4E6	206.709	467.330m#
2) T	3,5-DICHL...	6.290	5.859	702.8E6	265.5E6	326.692	300.135
3) T	4-Nitroph...	6.878	6.279	617.4E6	38434057	764.409	125.388 #
5) T	DICAMBA	7.240	6.801	1916.2E6	775.7E6	306.869	295.065
6) T	MCPD	7.409	6.905	64150876	58316168	13.740	27.112 #
7) T	MCPA	7.529	7.106	34510778	106.0E6	5.173	35.528 #
8) T	DICHLORPROP	7.927	7.413	572.1E6	244.2E6	363.720	365.858
9) T	2,4-D	8.154	7.692	737.7E6	243.6E6	405.101	331.047m
10) T	Pentachlo...	8.454	8.116	8157.4E6	3071.3E6	386.129	364.656
11) T	2,4,5-TP ...	9.004	8.477	2930.2E6	1193.2E6	346.268	334.726
12) T	2,4,5-T	9.290	8.875	2368.6E6	1513.7E6	267.217	462.776 #
13) T	2,4-DB	9.822	9.368	2308.7E6	777.8E6	1396.666	1757.857 #
14) T	DINOSEB	10.950	9.691	4408.2E6	755.9E6	671.271	302.968 #
15) T	Picloram	10.781	10.670	36174.8E6	6408.6E6	3030.754	1370.486 #
16) T	DCPA	11.250	10.605	4262.4E6	1702.9E6	394.491m	385.673

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011122\
Data File : PS017957.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2022 11:27
Operator : AJ\MA
Sample : N1052-01MS
Misc : LOT#M03961
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BALL-1MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/12/2022
Supervised By : Ankita Jodhani 01/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 15:29:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

