

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011124\
 Data File : PS025502.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 04:26
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 04:58:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 11 16:45:24 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	6.351	6.761	777.3E6	410.8E6	521.891	535.072
Target Compounds						
1) T Dalapon	0.000	2.558	0	3962759	N.D.	3.373
7) T MCPA	0.000	7.154f	0	9326657	N.D.	2.692
8) T DICHLORPROP	0.000	7.555	0	3295795	N.D.	3.813
12) T 2,4,5-T	0.000	8.992	0	838070	N.D.	<MDL
13) T 2,4-DB	0.000	9.489	0	430878	N.D.	<MDL
14) T DINOSEB	0.000	9.795	0	539780	N.D.	<MDL
15) T Picloram	0.000	10.718	0	661661	N.D.	<MDL
16) T DCPA	0.000	10.718	0	661661	N.D.	<MDL



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011124\
Data File : PS025502.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 04:26
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 04:58:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
Quant Title : 8080.M
QLast Update : Thu Jan 11 16:45:24 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 2.474 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
I.BLK




#1 Dalapon

R.T.: 2.558 min
Delta R.T.: 0.029 min
Response: 3962759
Conc: 3.37 ng/ml




#4 2,4-DCAA

R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 777338266
Conc: 521.89 ng/ml



#4 2,4-DCAA


R.T.: 6.761 min
Delta R.T.: 0.000 min
Response: 410828665
Conc: 535.07 ng/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 6.762 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
I.BLK




#7 MCPA

R.T.: 7.154 min
Delta R.T.: -0.067 min
Response: 9326657
Conc: 2.69 ug/ml



#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T.: 7.077 min
Response: 0
Conc: N.D.



#8 DICHLORPROP

R.T.: 7.555 min
Delta R.T.: 0.023 min
Response: 3295795
Conc: 3.81 ng/ml