

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011420\
 Data File : PS008415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2020 20:11
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

mohammad
 1/15/2020 4:19:13 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 23:49:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.271	5.970	1509.6E6	684.4E6	854.980m	722.306
Target Compounds							
1) T	Dalapon	2.536	2.228	1801.8E6	778.0E6	702.774	613.724
2) T	3,5-DICHL...	5.491	5.127	1871.1E6	943.3E6	781.178	675.138
3) T	4-Nitroph...	6.106	5.577	718.9E6	379.1E6	738.163	638.447
5) T	DICAMBA	6.465	6.140	5494.8E6	2755.9E6	799.515	689.151
6) T	MCPD	6.633	6.236	431.2E6	209.9E6	84.603m	69.014
7) T	MCPA	6.782	6.451	592.7E6	286.9E6	81.108	67.192
8) T	DICHLORPROP	7.154	6.764	1307.8E6	654.9E6	797.176	677.827
9) T	2,4-D	7.381	7.057	1422.4E6	792.4E6	781.911	675.953
10) T	Pentachlo...	7.697	7.493	18834.0E6	9272.4E6	774.719	686.572
11) T	2,4,5-TP ...	8.243	7.852	7382.8E6	3606.9E6	790.555	690.107
12) T	2,4,5-T	8.526	8.231	7207.5E6	3638.7E6	781.167	688.088
13) T	2,4-DB	9.079	8.738	1089.5E6	482.6E6	785.918	681.784
14) T	DINOSEB	10.244	9.075	5324.7E6	2518.9E6	742.132	666.780
15) T	Picloram	10.056	10.056	9140.0E6	4991.1E6	732.271	671.812
16) T	DCPA	10.509	10.006	9182.2E6	4564.4E6	773.708m	700.423

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011420\
Data File : PS008415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2020 20:11
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

mohammad
1/15/2020 4:19:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 23:49:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
Quant Title : 8080.M
QLast Update : Wed Jan 08 11:00:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

