

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011421\
 Data File : PS013750.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2021 10:32
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 00:23:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:51:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.207	6.909	781.3E6	295.1E6	817.572	733.885
Target Compounds							
1) T	Dalapon	3.029	2.622	794.9E6	344.5E6	730.832	662.563
2) T	3,5-DICHL...	6.424	6.064	997.1E6	431.1E6	754.003	684.334
3) T	4-Nitroph...	7.042	6.524	344.2E6	170.7E6	723.929	673.195
5) T	DICAMBA	7.396	7.075	3011.9E6	1213.7E6	782.108	706.703
6) T	MCPD	7.563	7.170	224.2E6	98713738	78.252	69.854
7) T	MCPA	7.712	7.384	312.7E6	138.5E6	76.082	68.502
8) T	DICHLORPROP	8.087	7.700	746.9E6	309.8E6	771.654	695.728
9) T	2,4-D	8.320	7.999	847.9E6	383.4E6	800.406	691.606
10) T	Pentachlo...	8.625	8.422	10890.1E6	4076.2E6	768.681	705.137
11) T	2,4,5-TP ...	9.174	8.784	4370.7E6	1701.0E6	784.895	712.102
12) T	2,4,5-T	9.462	9.165	4383.8E6	1623.9E6	775.258	709.348
13) T	2,4-DB	10.017	9.672	746.2E6	212.1E6	763.634	697.421
14) T	DINOSEB	11.166	9.998	3419.9E6	1191.5E6	775.150	712.045
15) T	Picloram	11.004	10.996	5482.1E6	2341.4E6	759.017	673.693
16) T	DCPA	11.432	10.925	5553.4E6	2215.5E6	786.021	715.839

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011421\
Data File : PS013750.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2021 10:32
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 00:23:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:51:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

