

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011421\
 Data File : PS013764.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2021 18:21
 Operator : DD\AJ
 Sample : PB134144BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB134144BS

Manual Integrations
 APPROVED

Ankita
 1/15/2021 1:22:47 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 00:28:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:51:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.206	6.908	576.4E6	237.1E6	603.237	589.597m
Target Compounds							
1) T	Dalapon	3.024	2.619	588.4E6	345.5E6	540.997	664.434m
2) T	3,5-DICHL...	6.422	6.064	770.0E6	295.6E6	582.257	469.218
3) T	4-Nitroph...	7.040	6.522	250.2E6	113.7E6	526.169	448.297
5) T	DICAMBA	7.395	7.075	2066.5E6	851.6E6	536.594	495.882
6) T	MCPD	7.561	7.168	168.3E6	70569326	58.740	49.938
7) T	MCPA	7.710	7.382	209.6E6	121.8E6	51.004	60.226
8) T	DICHLORPROP	8.086	7.700	532.2E6	227.1E6	549.819	510.031
9) T	2,4-D	8.318	7.997	630.6E6	298.9E6	595.304	539.069
10) T	Pentachlo...	8.624	8.421	8340.7E6	3165.5E6	588.730	547.603
11) T	2,4,5-TP ...	9.172	8.781	3307.5E6	1293.0E6	593.961	541.283
12) T	2,4,5-T	9.459	9.163	3264.7E6	1227.3E6	577.341	536.089
13) T	2,4-DB	10.012	9.669	400.9E6	155.1E6	410.329	510.090
14) T	DINOSEB	11.166	9.998	2444.1E6	878.3E6	553.993	524.868
15) T	Picloram	11.000	10.992	3533.9E6	1561.2E6	489.277	449.196
16) T	DCPA	11.431	10.924	3947.2E6	1656.4E6	558.681	535.165

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011421\
Data File : PS013764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2021 18:21
Operator : DD\AJ
Sample : PB134144BS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB134144BS

Manual Integrations
APPROVED

Ankita
1/15/2021 1:22:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 00:28:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:51:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

