

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011422\
 Data File : PS017988.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2022 10:54
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 18:48:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.048	6.645	639.0E6	296.0E6	442.046	510.202
Target Compounds							
9) T	2,4-D	8.212f	0.000	17387173	0	9.548	N.D. #
10) T	Pentachlo...	8.448	0.000	4968648	0	<MDL	N.D. #
12) T	2,4,5-T	9.290	8.924f	8979263	12158631	1.013	3.717 #
13) T	2,4-DB	9.842	0.000	8530730	0	5.161	N.D. #
14) T	DINOSEB	10.975	0.000	4634231	0	<MDL	N.D. #
15) T	Picloram	10.810	10.671	223.8E6	80255109	18.749	17.163
16) T	DCPA	11.244	0.000	3957462	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011422\
Data File : PS017988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2022 10:54
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 18:48:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.048 min
Delta R.T.: 0.000 min
Response: 638972668
Conc: 442.05 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.645 min
Delta R.T.: -0.006 min
Response: 296044003
Conc: 510.20 ng/ml




#9 2,4-D

R.T.: 8.212 min
Delta R.T.: 0.063 min
Response: 17387173
Conc: 9.55 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 7.710 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 9.290 min
Delta R.T.: 0.012 min
Response: 8979263
Conc: 1.01 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 8.924 min
Delta R.T.: 0.067 min
Response: 12158631
Conc: 3.72 ng/ml




#13 2,4-DB

R.T.: 9.842 min
Delta R.T.: 0.015 min
Response: 8530730
Conc: 5.16 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T. : 9.360 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.810 min
Delta R.T.: 0.014 min
Response: 223788432
Conc: 18.75 ng/ml

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 10.671 min
Delta R.T.: 0.012 min
Response: 80255109
Conc: 17.16 ng/ml